RHYTHM HEALTHCARE

TOC3 QUICK START GUIDE
TOC3 GUÍA DE INICIO



Patient Education Center
Scan the QR code to access product manuals, maintenance information, videos and other helpful resources.

Centro de Educación para el Paciente
Escanee el código QR para acceder a los manuales del producto, información de mantenimiento y otros recursos útiles.




Now that you have unpacked your TOC3, you are ready to begin setup. Setup is quick and simple; however, please read the entire user manual before using your TOC3 for the first time.







LANGUAGE SELECTION (ENGLISH OR SPANISH)
DEFAULT LANGUAGE IS ENGLISH



1. Press Power button to turn on TOC3.
2. Press and hold the Settings button. The default language is English.
3. Press and hold the Settings button again for 5 seconds.
   The interface will display: Switch to Spanish?
4. Then press the - or + to select Yes/No.
5. Press the Settings button to confirm Yes/No.
   - Confirm “Yes” to switch to Spanish
   - Confirm “No” to confirm English.
6. Press the Settings button.
   Return to the device startup interface, then you can use TOC3 normally.

CHECK FILTER REMINDER

When accumulated runtime has reached 600 hours, the Check Filter screen will appear.
Scan the QR code for instructions.
Press left arrow to check filter later.
Press right arrow if filter has already been checked and cleaned.
Device will then enter startup process and reset runtime.


Note: If you do nothing, the Check Filter screen will appear the next time you power on device.



STARTUP ERROR ALARM

An error will be displayed after startup if alarm persists more than three times consecutively during each startup.

Press left arrow to check later.
Press the right arrow if error has already been checked.
Device will then enter startup process.



QUICK SETUP INSTRUCTIONS

1



Find a well ventilated location to place the TOC3. Ensure the intake and exhaust have clear access.

2



Slide the flat side of the battery until the latch returns to the upper position and you hear it click into place.

3



Plug the AC Power Cord into the nearest electrical outlet.

4



Connect the other end of the AC Power Cord to the TOC3 power port.

5



Connect the Nasal Cannula to the Nozzle Fitting at the top of the TOC3. Ensure it is routed to prevent it from being pinched or kinked to avoid disruption of oxygen flow.

QUICK SETUP INSTRUCTIONS - CONTINUED

6



Press the On/Off button.
The device will beep and the indicator light will flash. A five minute warm-up time will initiate.

7



Wear the Nasal Cannula and breathe through your nose.

8



To Select the Flow Setting Prescribed by your Provider:
Press the Settings button to switch between Pulse and Continuous Modes. Press the + or - buttons to adjust the flow setting.

9



To Shut Down the TOC3:
Press and hold the On/Off button for three seconds.


Additional information:
When the adapter is connected or disconnected, or the battery is installed or removed, the device will beep and display the battery charge status.



Ahora que ha desempacado su TOC3, está listo para comenzar la configuración. La configuración es rápida y sencilla; sin embargo, lea todo el manual del usuario antes de utilizar su TOC3 por primera vez.



SELECCIÓN DE IDIOMAS (INGLÉS O ESPAÑOL)



1. Presione el botón de encendido para encender el TOC3.
2. Mantén presionado el botón de Configuración. El idioma predeterminado es inglés.
3. Presione nuevamente el botón de configuración durante 5 segundos. La pantalla mostrará: ¿Cambiar al español?
4. Luego, presione - o + y elija Sí/No.
5. Presione el botón de configuración para confirmar Sí/No.
   - Confirme “Sí” para cambiar al español.
   - Establezca “No” como inglés.
6. Presione el botón de configuración.
   Volviendo a la pantalla de arranque del dispositivo, se puede usar TOC3.

COMPRUEBE EL RECORDATORIO DEL FILTRO

Cuando el tiempo acumulado de funcionamiento haya alcanzado las 600 horas, aparecerá la pantalla de Verificación del Filtro.
Escanee el código QR para obtener instrucciones.
Presione la flecha izquierda para verificar el filtro más tarde.
Presione la flecha derecha si el filtro ya ha sido verificado y limpiado.
El dispositivo entonces entrará en el proceso de arranque y reiniciará el tiempo de funcionamiento.

Nota: Si no realiza ninguna acción, la pantalla de Verificación del Filtro aparecerá la próxima vez que encienda el dispositivo.



ALARMA DE ERROR DE INICIO

Se mostrará un error después de iniciar si la alarma persiste más de tres veces consecutivas durante cada inicio.

Pulse la flecha izquierda para revisar más tarde.
Pulse la flecha derecha si el error ya ha sido verificado.
El dispositivo comenzará el proceso de inicio.



INSTRUCCIONES DE INSTALACIÓN RÁPIDA

1



Busque un lugar bien ventilado para colocar el TOC3. Asegúrese de mantener despejadas la entrada de aire y puerto de salida de oxígeno.

2



Deslice la bateria horizontalmente hacia su lugar hasta que el pestillo vuelva a la posición superior junto con un sonido.

3



Enchufe la fuente de alimentación CA en el tomacorriente más cercano.

4



Conecte el otro extremo del cable de alimentación de CA al puerto de energía del TOC3.

5



Coloque la cánula nasal hasta el fondo del puerto de salida de oxígeno.
Asegúrese que la cánula no se doble o estrangule así se evita la interrupción de salida de oxigeno.

INSTRUCCIONES DE INSTALACIÓN RÁPIDA - CONTINUACIÓN

6



Presione el botón ENCENDIDO/ APAGADO, se escuchará un pitido y la luz se encenderá. Se iniciará un tiempo de calentamiento de 2 minutos.

7



Coloque la cánula nasal en su rostro y respire a través de su nariz.

8



Para Seleccionar el Ajuste de Flujo Prescrito por su Proveedor:
Pulse el botón de configuración para alternar entre el modo de pulso y el modo continuo. Presione los botones + o - para ajustar al nivel deseado, ya sea subir o bajar.

9



Para Apagar el TOC3:
Presione el botón ENCENDIDO/ APAGADO durante 3 segundos.

Información Adicional:
Cuando se conecta o desconecta el adaptador, o se instala o retira la batería, el dispositivo emitirá un pitido y mostrará el estado de carga de la batería.

RHYTHM HEALTHCARE

Rhythm Healthcare, LLC
489 Fifth Avenue, 28th FL
New York, NY 10017

contactus@rhythmhc.com
www.rhythmhc.com
(877) 843-6464

RHC-TOC-003, V5: 8/4/2026